

表 399 ETRF装着率(%) 細菌数(cfu/mL)別（ベッドサイドコンソール台数≧1）

細菌数	0(装着せず)	10未満	10～	20～	30～	40～	50～	60～	70～	80～	90～	100(全台装着)	合計	記載なし	総計	平均	標準偏差
0.1未満	28	29	22	30	26	17	28	35	39	63	125	2,367	2,809	2	2,811	93.80	19.60
(%)	(1.0)	(1.0)	(0.8)	(1.1)	(0.9)	(0.6)	(1.0)	(1.2)	(1.4)	(2.2)	(4.4)	(84.3)	(100.0)				
0.1～	19	10	16	11	10	6	11	9	11	15	35	397	550	1	551	86.20	29.21
(%)	(3.5)	(1.8)	(2.9)	(2.0)	(1.8)	(1.1)	(2.0)	(1.6)	(2.0)	(2.7)	(6.4)	(72.2)	(100.0)				
1～	21	17	14	13	13	4	5	4	9	8	13	250	371	1	372	79.02	35.77
(%)	(5.7)	(4.6)	(3.8)	(3.5)	(3.5)	(1.1)	(1.3)	(1.1)	(2.4)	(2.2)	(3.5)	(67.4)	(100.0)				
10～	15	5	7	5	5	4	9	6	3	4	4	89	156		156	72.61	37.74
(%)	(9.6)	(3.2)	(4.5)	(3.2)	(3.2)	(2.6)	(5.8)	(3.8)	(1.9)	(2.6)	(2.6)	(57.1)	(100.0)				
100以上	2	1	2	2	2	3				1	1	26	40		40	76.87	36.23
(%)	(5.0)	(2.5)	(5.0)	(5.0)	(5.0)	(7.5)				(2.5)	(2.5)	(65.0)	(100.0)				
合計	85	62	61	61	56	34	53	54	62	91	178	3,129	3,926	4	3,930	90.33	24.93
(%)	(2.2)	(1.6)	(1.6)	(1.6)	(1.4)	(0.9)	(1.3)	(1.4)	(1.6)	(2.3)	(4.5)	(79.7)	(100.0)				
不明	54	7	7	8	4	2	9	6	6	5	9	145	262	2	264	67.28	42.67
(%)	(20.6)	(2.7)	(2.7)	(3.1)	(1.5)	(0.8)	(3.4)	(2.3)	(2.3)	(1.9)	(3.4)	(55.3)	(100.0)				
記載なし	20	4	4	2	2	1	1	2	1	1	1	65	104	6	110	69.12	43.26
(%)	(19.2)	(3.8)	(3.8)	(1.9)	(1.9)	(1.0)	(1.0)	(1.9)	(1.0)	(1.0)	(1.0)	(62.5)	(100.0)				
総計	159	73	72	71	62	37	63	62	69	97	188	3,339	4,292	12	4,304	88.41	27.64
(%)	(3.7)	(1.7)	(1.7)	(1.7)	(1.4)	(0.9)	(1.5)	(1.4)	(1.6)	(2.3)	(4.4)	(77.8)	(100.0)				

数値下のかっこ内は行方向の合計に対する%です。
(C)Japanese Society for Dialysis Therapy
2014年調査