

表 423 ET測定値(EU/mL) 細菌数測定の種類別 (ベッドサイドコンソール台数≧1)

細菌数測定の 頻度	0.001EU/mL未満	0.001～0.01EU/mL 未満	0.01～0.05EU/mL 未満	0.05～0.1EU/mL未 満	0.1～0.25EU/mL未 満	0.25～0.5EU/mL未 満	0.5EU/mL以上	合計	不明	記載なし	総計
なし	63	26	18	4	3		1	115	77	50	242
(%)	(54.8)	(22.6)	(15.7)	(3.5)	(2.6)		(0.9)	(100.0)			
毎日	11	4			1			16			16
(%)	(68.8)	(25.0)			(6.3)			(100.0)			
毎週	102	16	7	1	1			127	1		128
(%)	(80.3)	(12.6)	(5.5)	(0.8)	(0.8)			(100.0)			
隔週	209	29	9	2	6	1	3	259	1		260
(%)	(80.7)	(11.2)	(3.5)	(0.8)	(2.3)	(0.4)	(1.2)	(100.0)			
毎月	2,172	361	123	37	21	10	12	2,736	8		2,744
(%)	(79.4)	(13.2)	(4.5)	(1.4)	(0.8)	(0.4)	(0.4)	(100.0)			
年数回	303	58	28	9	11	4	1	414	8		422
(%)	(73.2)	(14.0)	(6.8)	(2.2)	(2.7)	(1.0)	(0.2)	(100.0)			
年1回	296	50	21	9	5	5	5	391	4		395
(%)	(75.7)	(12.8)	(5.4)	(2.3)	(1.3)	(1.3)	(1.3)	(100.0)			
合計	3,156	544	206	62	48	20	22	4,058	99	50	4,207
(%)	(77.8)	(13.4)	(5.1)	(1.5)	(1.2)	(0.5)	(0.5)	(100.0)			
不明	10	4	5	1				20	65	5	90
(%)	(50.0)	(20.0)	(25.0)	(5.0)				(100.0)			
記載なし	1							1		6	7
(%)	(100.0)							(100.0)			
総計	3,167	548	211	63	48	20	22	4,079	164	61	4,304
(%)	(77.6)	(13.4)	(5.2)	(1.5)	(1.2)	(0.5)	(0.5)	(100.0)			

数値下のかっこ内は行方向の合計に対する%です。

(C)Japanese Society for Dialysis Therapy

2014年調査